

checkCIF/PLATON report

Structure factors have been supplied for datablock(s) AX2866absneu

THIS REPORT IS FOR GUIDANCE ONLY. IF USED AS PART OF A REVIEW PROCEDURE FOR PUBLICATION, IT SHOULD NOT REPLACE THE EXPERTISE OF AN EXPERIENCED CRYSTALLOGRAPHIC REFEREE.

No syntax errors found. CIF dictionary Interpreting this report

Datablock: AX2866absneu

Bond precision: C-C = 0.0069 Å

Wavelength=1.54178

Cell: a=8.4394 (4) b=13.0868 (7) c=13.2386 (7)
 alpha=119.294 (3) beta=103.450 (3) gamma=91.119 (4)
Temperature: 150 K

	Calculated	Reported
Volume	1224.76 (12)	1224.76 (11)
Space group	P -1	P -1
Hall group	-P 1	-P 1
Moiety formula	C24 H28 N6 O8	C24 H28 N6 O8
Sum formula	C24 H28 N6 O8	C24 H28 N6 O8
Mr	528.52	528.52
Dx, g cm ⁻³	1.433	1.433
Z	2	2
Mu (mm ⁻¹)	0.922	0.922
F000	556.0	556.0
F000'	557.91	
h, k, lmax	9, 15, 15	9, 15, 15
Nref	3898	3821
Tmin, Tmax	0.931, 0.958	0.419, 0.520
Tmin'	0.665	

Correction method= # Reported T Limits: Tmin=0.419 Tmax=0.520
AbsCorr = MULTI-SCAN

Data completeness= 0.980

Theta (max)= 62.470

R(reflections)= 0.0730 (3177)

wR2(reflections)=
0.2392 (3821)

S = 1.097

Npar= 347

The following ALERTS were generated. Each ALERT has the format

test-name_ALERT_alert-type_alert-level.

Click on the hyperlinks for more details of the test.

Alert level C

THETM01_ALERT_3_C The value of sine(theta_max)/wavelength is less than 0.590
Calculated sin(theta_max)/wavelength = 0.5752

PLAT340_ALERT_3_C Low Bond Precision on C-C Bonds 0.00694 Ang.
PLAT906_ALERT_3_C Large K Value in the Analysis of Variance 6.780 Check
PLAT906_ALERT_3_C Large K Value in the Analysis of Variance 2.086 Check
PLAT911_ALERT_3_C Missing FCF Refl Between Thmin & STh/L= 0.575 77 Report

0	4	0,	-1	6	0,	-2	7	0,	-1	7	0,	-7	10	0,	7	-10	1,
1	-7	1,	3	-6	1,	2	-4	1,	5	-4	1,	1	-3	1,	-4	-2	1,
-1	-2	1,	1	-1	1,	-1	1	1,	3	2	1,	-1	6	1,	5	9	1,
1	-7	2,	2	-7	2,	0	-6	2,	2	-6	2,	3	-6	2,	2	-5	2,
1	-4	2,	5	-4	2,	1	-3	2,	2	-2	2,	1	-1	2,	-5	3	2,

(47 More Missing: see the .ckf listing file)

Alert level G

PLAT883_ALERT_1_G Absent Datum for _atom_sites_solution_primary .. Please Do !
PLAT909_ALERT_3_G Percentage of I>2sig(I) Data at Theta(Max) Still 70% Note
PLAT933_ALERT_2_G Number of HKL-OMIT Records in Embedded .res File 19 Note

2	-1	3,	0	-6	2,	1	-1	2,	1	-4	6,	2	2	3,	2	-2	4,
1	3	4,	-1	-2	1,	1	-5	6,	3	2	1,	4	-2	7,	2	-1	10,
0	2	3,	-1	-2	5,	1	-3	5,	5	-2	6,	2	-2	2,	1	-1	1,
0	-7	3,															

PLAT965_ALERT_2_G The SHELXL WEIGHT Optimisation has not Converged Please Check
PLAT967_ALERT_5_G Note: Two-Theta Cutoff Value in Embedded .res .. 125.0 Degree
PLAT969_ALERT_5_G The 'Henn et al.' R-Factor-gap value 5.620 Note
Predicted wR2: Based on SigI**2 4.26 or SHELX Weight 21.80
PLAT978_ALERT_2_G Number C-C Bonds with Positive Residual Density. 0 Info

- 0 **ALERT level A** = Most likely a serious problem - resolve or explain
0 **ALERT level B** = A potentially serious problem, consider carefully
5 **ALERT level C** = Check. Ensure it is not caused by an omission or oversight
7 **ALERT level G** = General information/check it is not something unexpected

- 1 ALERT type 1 CIF construction/syntax error, inconsistent or missing data
3 ALERT type 2 Indicator that the structure model may be wrong or deficient
6 ALERT type 3 Indicator that the structure quality may be low
0 ALERT type 4 Improvement, methodology, query or suggestion
2 ALERT type 5 Informative message, check
-
-

It is advisable to attempt to resolve as many as possible of the alerts in all categories. Often the minor alerts point to easily fixed oversights, errors and omissions in your CIF or refinement strategy, so attention to these fine details can be worthwhile. In order to resolve some of the more serious problems it may be necessary to carry out additional measurements or structure refinements. However, the purpose of your study may justify the reported deviations and the more serious of these should normally be commented upon in the discussion or experimental section of a paper or in the "special_details" fields of the CIF. checkCIF was carefully designed to identify outliers and unusual parameters, but every test has its limitations and alerts that are not important in a particular case may appear. Conversely, the absence of alerts does not guarantee there are no aspects of the results needing attention. It is up to the individual to critically assess their own results and, if necessary, seek expert advice.

Publication of your CIF in IUCr journals

A basic structural check has been run on your CIF. These basic checks will be run on all CIFs submitted for publication in IUCr journals (*Acta Crystallographica*, *Journal of Applied Crystallography*, *Journal of Synchrotron Radiation*); however, if you intend to submit to *Acta Crystallographica Section C* or *E* or *IUCrData*, you should make sure that full publication checks are run on the final version of your CIF prior to submission.

Publication of your CIF in other journals

Please refer to the *Notes for Authors* of the relevant journal for any special instructions relating to CIF submission.

