Online Resource 1

Determination of pesticide residues, mycotoxins and benzo(a)pyrene in infant foods: quantification and reliability of results
Accreditation and Quality Assurance – Babincev L.M., Kulizić J.M., Sikimić A.Ž.

(corresponding author: ljiljana.babincev@pr.ac.rs)

Redacted laboratory test reports for all 20 samples (official control, July–August 2023), followed by the laboratory's method legends (L1–L3). The name and address of the laboratory, product and manufacturer names, lot numbers, laboratory sample identifiers, stamps, approval signatures and names of analysts have been removed at the request of the body that commissioned the control; all analytical content (parameters, results, reference values, method codes) is reproduced unchanged. Values are expressed in mg/kg (pesticides) and µg/kg (mycotoxins, benzo(a)pyrene); "<" denotes a result below the limit of quantification. The reports were issued in Serbian; column headings: Parametar = parameter, Vrednost = result, Referentna vrednost = maximum level, Oznaka metode = method code.

Key to sample codes

Code	Product group	Intended use	Package	Analysis completed	Report pages
F1	Infant formula (from birth)	from birth	powder, 400 g	13 Jul 2023	10
F2	Infant formula (from birth)	from birth	powder, 800 g	13 Jul 2023	10
F3	Infant formula, FSMP anti-regurgitation	from birth	powder, 400 g	13 Jul 2023	10
F4	Infant formula, hypoallergenic	from birth	powder, 400 g	13 Jul 2023	10
F5	Infant formula, lactose-free	from birth	powder, 400 g	13 Jul 2023	10
F6	Follow-on formula	from 6 months	powder, 800 g	13 Jul 2023	10
F7	Follow-on formula (comfort)	from 6 months	powder, 800 g	13 Jul 2023	10
F8	Young-child formula with L. reuteri	from 12 months	powder, 800 g	13 Jul 2023	10
F9	Young-child formula with L. reuteri (comfort)	from 12 months	powder, 800 g	13 Jul 2023	10
D1	Milk dessert, vanilla	from 6 months	retail pack	13 Jul 2023	10
D2	Milk dessert, vanilla	from 6 months	retail pack	13 Jul 2023	10
D3	Milk dessert, semolina	from 6 months	retail pack	13 Jul 2023	10
D4	Milk dessert, banana	from 6 months	retail pack	13 Jul 2023	10
D5	Milk dessert, biscuit	from 6 months	retail pack	13 Jul 2023	10
D6	Milk dessert, biscuit	from 6 months	retail pack	13 Jul 2023	10
D7	Milk dessert, cocoa	from 8 months	retail pack	13 Jul 2023	10
D8	Milk dessert, cocoa	from 8 months	retail pack	13 Jul 2023	10
C1	Organic instant fruit cereal (apple, peach, mango)	from 6 months	250 g	14 Aug 2023	12
C2	Organic instant fruit cereal (apple, banana, wholegrain)	from 6 months	250 g	14 Aug 2023	12
C3	Organic instant fruit cereal (apple, banana, wholegrain)	from 6 months	250 g	14 Aug 2023	12
L1	Method legend – formulae				1
L2	Method legend – milk desserts				1
L3	Method legend – fruit cereals				1

Sample code: F1 (product name, lot and sample ID redacted)

C REZULTATI ISPITIVANJA

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA

Datum i vreme završetka ispitivanja: 13.07.2023. u 10:12

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)

Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroaniline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bupirimate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predvideni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethionfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (***GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenthion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluridone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, cis- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Heptachlor epoxide, trans- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodione (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton (pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)	< 0,003	< 0,006	**SRPS EN 15662:2018
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers Including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Molinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthiuron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predvideni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Valifenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dithiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin M1	< 0,01	< 0,025	VDM 0244
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 31 i čl. 32 Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

LABORATORY REPORT – sample F2 – page 1/10 (header redacted)

Sample code: F2 (product name, lot and sample ID redacted)

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA
Datum i vreme završetka ispitivanja: 13.07.2023. u 10:14

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroaniline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bupirimate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethionfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (***GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenthion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluridone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, cis- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Heptachlor epoxide, trans- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodione (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton (pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)	< 0,003	< 0,006	**SRPS EN 15662:2018
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Molinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthiuron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Valifenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dithiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin M1	< 0,01	< 0,025	VDM 0244
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 31 i čl. 32 Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

Sample code: F3 (product name, lot and sample ID redacted)

C REZULTATI ISPITIVANJA

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA

Datum i vreme završetka ispitivanja: 13.07.2023. u 10:24

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)

Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroaniline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bupirimate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethionfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (***GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenthion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluridone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, cis- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Heptachlor epoxide, trans- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodione (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton (pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)	< 0,003	< 0,006	**SRPS EN 15662:2018
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers Including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Molinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predvideni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthiuron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Valifenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dithiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin M1	< 0,01	< 0,025	VDM 0244
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 31 i čl. 32 Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

LABORATORY REPORT – sample F4 – page 1/10 (header redacted)

Sample code: F4 (product name, lot and sample ID redacted)

C REZULTATI ISPITIVANJA

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA

Datum i vreme završetka ispitivanja: 13.07.2023. u 10:31

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)

Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroaniline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bupirimate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predvideni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethionfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (***GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenthion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluridone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, cis- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Heptachlor epoxide, trans- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodione (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton (pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)	< 0,003	< 0,006	**SRPS EN 15662:2018
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Molinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthiuron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Valifenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dithiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin M1	< 0,01	< 0,025	VDM 0244
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 31 i čl. 32 Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

Sample code: F5 (product name, lot and sample ID redacted)

C REZULTATI ISPITIVANJA

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA

Datum i vreme završetka ispitivanja: 13.07.2023. u 10:45

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dječetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)

Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroanilline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predvideni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bupirimate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethionfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (***GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenthion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluridone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, cis- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Heptachlor epoxide, trans- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodione (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton (pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)	< 0,003	< 0,006	**SRPS EN 15662:2018
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Molinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predvideni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthiuron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

LABORATORY REPORT – sample F5 – page 10/10 (header redacted)

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Valifenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dithiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin M1	< 0,01	< 0,025	VDM 0244
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 31 i čl. 32 Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

Parametar	Vrednost	Referentna vrednost	Oznaka metode
Sadržaj laktoze, %	< 1		****Pravilnik[11] Metoda 13

LABORATORY REPORT – sample F6 – page 1/10 (header redacted)

Sample code: F6 (product name, lot and sample ID redacted)

C REZULTATI ISPITIVANJA

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA

Datum i vreme završetka ispitivanja: 13.07.2023. u 09:56

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dječetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)

Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroaniline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprirmate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predvideni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethionfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (***GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fonthion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluridone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, cis- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predvideni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Heptachlor epoxide, trans- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodione (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton (pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)	< 0,003	< 0,006	**SRPS EN 15662:2018
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mollnate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanll (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthiuron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Valifenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dithiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin M1	< 0,01	< 0,025	VDM 0244
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 31 i čl. 32 Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

Sample code: F7 (product name, lot and sample ID redacted)

C REZULTATI ISPITIVANJA

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA

Datum i vreme završetka ispitivanja: 13.07.2023. u 10:04

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroaniline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprimate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethionfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (**GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenthion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluridone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, cis-	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
(GC-MS/MS)			
Heptachlor epoxide, trans- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodione (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton (pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)	< 0,003	< 0,006	**SRPS EN 15662:2018
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Molinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthiuron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Valifenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dithiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin M1	< 0,01	< 0,025	VDM 0244
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 31 i čl. 32 Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

LABORATORY REPORT – sample F8 – page 1/10 (header redacted)

Sample code: F8 (product name, lot and sample ID redacted)

C REZULTATI ISPITIVANJA

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA

Datum i vreme završetka ispitivanja: 13.07.2023. u 10:09

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroaniline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprimate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethionfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (***GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenthion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluridone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, cis- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Heptachlor epoxide, trans- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodione (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton (pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)	< 0,003	< 0,006	**SRPS EN 15662:2018
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predvideni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Molinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthiuron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vallfenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dithiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin M1	< 0,01	< 0,025	VDM 0244
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 31 i čl. 32 Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

LABORATORY REPORT – sample F9 – page 1/10 (header redacted)

Sample code: F9 (product name, lot and sample ID redacted)

C REZULTATI ISPITIVANJA

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA

Datum i vreme završetka ispitivanja: 13.07.2023. u 10:33

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022))

Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroanlline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022))			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bupirimate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022))			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predvideni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022))			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethlonfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (***GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenthion (CC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predvideni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022))			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluridone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, cis- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022))			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Heptachlor epoxide, trans- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodione (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton (pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)	< 0,003	< 0,006	**SRPS EN 15662:2018
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022))			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Molinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022))			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022))			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthiuron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022))			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Valifenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dithiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin M1	< 0,01	< 0,025	VDM 0244
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 31 i čl. 32 Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

(approval, signatures and names redacted)

LABORATORY REPORT – sample D1 – page 1/10 (header redacted)

Sample code: D1 (product name, lot and sample ID redacted)

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA

Datum i vreme završetka ispitivanja: 13.07.2023. u 09:11

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)

Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroaniline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predvideni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bupirimate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethionfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (***GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenthion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flurldone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, cis- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, trans-	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
(GC-MS/MS)			
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodione (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton (pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)	< 0,003	< 0,006	**SRPS EN 15662:2018
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Molinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthiuron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Valifenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dithiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin B1	< 0,05	< 0,1	VDM 0013
Ohratoksin A	< 0,25	< 0,5	VDM 0014
Zearalenon	< 5	< 20	VDM 0227
Deoksinivalenol	< 25	< 200	VDM 0246
Fumonizini (zbir B1 i B2)	< 25	< 200	VDM 0289
Aflatoksin M1	< 0,01		VDM 0244
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 50 i čl. 51. Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

Sample code: D2 (product name, lot and sample ID redacted)

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA
Datum i vreme završetka ispitivanja: 13.07.2023. u 09:21

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dječetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroanilline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buplrimate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethionfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (***GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenthion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluridone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, cis- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, trans-	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
(GC-MS/MS)			
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodione (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton (pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)	< 0,003	< 0,006	**SRPS EN 15662:2018
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Molinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthiuron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vallfenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dithiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin B1	< 0,05	< 0,1	VDM 0013
Ohratoksin A	< 0,25	< 0,5	VDM 0014
Zearalenon	< 5	< 20	VDM 0227
Deoksinivalenol	< 25	< 200	VDM 0246
Fumonizini (zbir B1 i B2)	< 25	< 200	VDM 0289
Aflatoksin M1	< 0,01		VDM 0244
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 50 i čl. 51. Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

Sample code: D3 (product name, lot and sample ID redacted)

C REZULTATI ISPITIVANJA

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA

Datum i vreme završetka ispitivanja: 13.07.2023. u 09:24

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroaniline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
(LC-MS/MS)			
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bupirimate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethionfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (***GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Fenthion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluridone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Heptachlor epoxide, cis- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, trans- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodione (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton (pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)	< 0,003	< 0,006	**SRPS EN 15662:2018
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
(GC-MS/MS;LC-MS/MS)			
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Molinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthiuron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Valifenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dithiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin B1	< 0,05	< 0,1	VDM 0013
Ohratoksin A	< 0,25	< 0,5	VDM 0014
Zearalenon	< 5	< 20	VDM 0227
Deoksinivalenol	< 25	< 200	VDM 0246
Fumonizini (zbir B1 i B2)	< 25	< 200	VDM 0289
Aflatoksin M1	< 0,01		VDM 0244
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 50 i čl. 51. Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

Sample code: D4 (product name, lot and sample ID redacted)

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA
Datum i vreme završetka ispitivanja: 13.07.2023. u 09:28

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dječetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroaniline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buplrimate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predvideni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethionfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (***GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenthion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluridone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, cis- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, trans-	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
(GC-MS/MS)			
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodione (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton (pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)	< 0,003	< 0,006	**SRPS EN 15662:2018
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Molinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthiuron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Valifenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dithiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin B1	< 0,05	< 0,1	VDM 0013
Ohratoksin A	< 0,25	< 0,5	VDM 0014
Zearalenon	< 5	< 20	VDM 0227
Deoksinivalenol	< 25	< 200	VDM 0246
Fumonizini (zbir B1 i B2)	< 25	< 200	VDM 0289
Aflatoksin M1	< 0,01		VDM 0244
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 50 i čl. 51. Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

Sample code: D5 (product name, lot and sample ID redacted)

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA
Datum i vreme završetka ispitivanja: 13.07.2023. u 09:30

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dječetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroanilline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bupirimate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethionfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (***GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenthion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluridone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, cis- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, trans-	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
(GC-MS/MS)			
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodlone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton (pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)	< 0,003	< 0,006	**SRPS EN 15662:2018
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Molinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthiuron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vallfenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dithiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin B1	< 0,05	< 0,1	VDM 0013
Ohratoksin A	< 0,25	< 0,5	VDM 0014
Zearalenon	< 5	< 20	VDM 0227
Deoksinivalenol	< 25	< 200	VDM 0246
Fumonizini (zbir B1 i B2)	< 25	< 200	VDM 0289
Aflatoksin M1	< 0,01		VDM 0244
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 50 i čl. 51. Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

Sample code: D6 (product name, lot and sample ID redacted)

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA
Datum i vreme završetka ispitivanja: 13.07.2023. u 09:42

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroaniline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bupirimate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethlon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethionfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (***GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenthion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluridone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, cis- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, trans-	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
(GC-MS/MS)			
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodione (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton (pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)	< 0,003	< 0,006	**SRPS EN 15662:2018
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predvideni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Molinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthiuron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Valifenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dithiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin B1	< 0,05	< 0,1	VDM 0013
Ohratoksin A	< 0,25	< 0,5	VDM 0014
Zearalenon	< 5	< 20	VDM 0227
Deoksinivalenol	< 25	< 200	VDM 0246
Fumonizini (zbir B1 i B2)	< 25	< 200	VDM 0289
Aflatoksin M1	< 0,01		VDM 0244
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 50 i čl. 51. Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

Sample code: D7 (product name, lot and sample ID redacted)

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA
Datum i vreme završetka ispitivanja: 13.07.2023. u 09:45

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroaniline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bupirimate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predvideni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predvideni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethionfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (***GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenthion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluridone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, cis- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, trans-	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
(GC-MS/MS)			
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodione (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton (pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)	< 0,003	< 0,006	**SRPS EN 15662:2018
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mollnate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predvideni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthiuron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Valifenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dithiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin B1	< 0,05	< 0,1	VDM 0013
Ohratoksin A	< 0,25	< 0,5	VDM 0014
Zearalenon	< 5	< 20	VDM 0227
Deoksinivalenol	< 25	< 200	VDM 0246
Fumonizini (zbir B1 i B2)	< 25	< 200	VDM 0289
Aflatoksin M1	< 0,01		VDM 0244
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 50 i čl. 51. Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

Sample code: D8 (product name, lot and sample ID redacted)

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA
Datum i vreme završetka ispitivanja: 13.07.2023. u 09:53

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroaniline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl (LC MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bupirimate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethionfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (***GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenthion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluridone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, cis- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Heptachlor epoxide, trans- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodione (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton (pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)	< 0,003	< 0,006	**SRPS EN 15662:2018
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Molinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthluron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Valifenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dithiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin B1	< 0,05	< 0,1	VDM 0013
Ohratoksin A	< 0,25	< 0,5	VDM 0014
Zearalenon	< 5	< 20	VDM 0227
Deoksinivalenol	< 25	< 200	VDM 0246
Fumonizini (zbir B1 i B2)	< 25	< 200	VDM 0289
Aflatoksin M1	< 0,01		VDM 0244
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 50 i čl. 51. Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

(approval, signatures and names redacted)

LABORATORY REPORT – sample C1 – page 1/12 (header redacted)

Sample code: C1 (product name, lot and sample ID redacted)

C REZULTATI ISPITIVANJA

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA

Datum i vreme završetka ispitivanja: 14.08.2023. u 13:58

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dječetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroaniline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bupirimate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethionfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (***GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenthion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluridone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, cis- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Heptachlor epoxide, trans- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodione (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton	< 0,003	< 0,006	**SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
(pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)			
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Molinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predvideni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans- isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamlfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthiuron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Valifenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dithiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin B1	< 0,05	< 0,1	VDM 0013
Ohratoksin A	< 0,25	< 0,5	VDM 0014
Zearalenon	< 5	< 20	VDM 0227
Deoksinivalenol	< 25	< 200	VDM 0246
Patulin	< 4	< 10	VDM 0015
Fumonizini (zbir B1 i B2)	< 25	< 200	VDM 0289
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 50 i čl. 51 Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

Sample code: C2 (product name, lot and sample ID redacted)

C REZULTATI ISPITIVANJA

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA

Datum i vreme završetka ispitivanja: 14.08.2023. u 14:10

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroaniline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bupirimate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethionfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (***GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenthion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluridone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, cis- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, trans- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodione (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton (pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)	< 0,003	< 0,006	**SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Molinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans- isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthiuron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
(GC-MS/MS;LC-MS/MS)			
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Valifenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dithiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin B1	< 0,05	< 0,1	VDM 0013
Ohratoksin A	< 0,25	< 0,5	VDM 0014
Zearalenon	< 5	< 20	VDM 0227
Deoksinivalenol	< 25	< 200	VDM 0246
Patulin	< 4	< 10	VDM 0015
Fumonizini (zbir B1 i B2)	< 25	< 200	VDM 0289
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 50 i čl. 51 Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

Sample code: C3 (product name, lot and sample ID redacted)

C REZULTATI ISPITIVANJA

REZULTATI FIZIČKIH, FIZIČKO-HEMIJSKIH I HEMIJSKIH ISPITIVANJA

Datum i vreme završetka ispitivanja: 14.08.2023. u 14:13

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)

Parametar	Vrednost	Referentna vrednost	Oznaka metode
PESTICIDI mg/kg			
2,4-D (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
2-Phenylphenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3,5 Dichloroaniline (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
3 - Hidroxycarbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Abamectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetamiprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acephate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acetochlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aclonifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Acrinathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Alachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb sulfoxide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Aldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Allethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametoctradin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ametryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amidosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Amitraz (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Antraquinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Atrazine-desethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azinphos-methyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Azoxystrobin (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predvideni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Benalaxyl (GC-MS/MS; LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bendiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Benthiavalicarb-isopropyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bifenthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bioallethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Biphenyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bitertanol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Boscalid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromophos- methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromopropylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bromuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Bupirimate (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Buprofezin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butafenacil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Butylate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cadusafos (GC-MS/MS)	< 0,003	< 0,006	SRPS EN 15662:2018
Captafol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Captan (sum of captan and THPI, expressed as captan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbaryl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbendazim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbofuran (LC-MS/MS)	< 0,001	< 0,01	SRPS EN 15662:2018
Carbophenothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Carbosulfan (LC-MS/MS)	< 0,002	< 0,01	SRPS EN 15662:2018
Carboxin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorantraniliprole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlordane, cis- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chordane, trans- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenapyr (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfenvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorfluazuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Chlorobenzilate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorothalonil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	VDM 0299
Chlorotoluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chloroxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpropham (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorpyrifos - methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Chlorthal-dimethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clofentezine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clomazone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Clothianidin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Coumaphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyanazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyazofamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cycloate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyflufenamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyfluthrin, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lambda-cyhalothrin (includes gamma-cyhalothrin) (sum of R,S and S,R isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cymoxanil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cypermethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyproconazole 1&2 (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyprodinil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Cyromazine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDD, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDE, p, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
DDT, o, p` (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Deltamethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diazinon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichloran (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Dichlorbenil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dichlorvos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diclobutrazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dicofol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dieldrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Diethofencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Difenoconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethoate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimethomorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dimoxystrobin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diniconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dioxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenamid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Diphenylamine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Disulfoton (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Disulfoton sulfoxide (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Dithianon (LC-MS/MS)	< 0,01	< 0,01	VDM 0297
Diuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dodine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Emmamectin benzoate B1a (expressed as Emmamectin) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endosulfan-sulphate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Endrin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Epoxiconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
EPTC (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Esfenvalerate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etaconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethionfencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethirimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ethoprophos (GC-MS/MS)	< 0,003	< 0,008	SRPS EN 15662:2018
Etofenprox (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Etoxazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etridiazole (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Etrimfos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Famoxadone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamidone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenamiphos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenarimol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenazaquin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenbuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenchlorphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenhexamid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenitrothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenobucarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenoxycarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropathrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropidin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpropimorph (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenpyroximate (E,Z) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fensulfothion (***GC-MS/MS;LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion oxon sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fensulfothion sulfone (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenthion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fentin (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Fenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fenvalerate (sum of isomers RR, SS, RS & SR) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fipronil (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-desulfinyl (GC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Fipronil-sulfone (GC-MS/MS;LC-MS/MS)	< 0,003	< 0,004	SRPS EN 15662:2018
Flonicamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluazifop-p-butyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Fluazinam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flubendiamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flucythrinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fludioxonil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flufenoxuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluometuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopicolide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluopyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluoxastrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluquinconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluridone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluroxypyr-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flusilazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutolanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Flutriafol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fluxapyroxad (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Folpet (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fonofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron	< 0,01	< 0,01	SRPS EN 15662:2018
Foramsulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Forchlorfenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formetanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Formothion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fosthiazate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Fuberidazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Furalaxyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Glyphosate (LC-MS/MS)	< 0,01	< 0,01	VDM 0295
Haloxyfop-2-ethoxyethyl (GC-MS/MS)	< 0,003	< 0,003	***SRPS EN 15662:2018
HCH, α- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, β- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, γ- (Lindan) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
HCH, δ- (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Heptachlor (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, cis- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Heptachlor epoxide, trans- (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Heptenophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexachlorobenzene (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Hexaconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexaflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexazinone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Hexythiazox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazalil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imazamox (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Imidacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Indoxacarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iodosulfuron-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ipconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprodione (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Iprovalicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isocarbophos (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoprocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoproturon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Isoxaflutole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Ivermectin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Kresoxim-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Linuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Lufenuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malaoxon (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Malathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mandipropamid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepanipyrim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mepronil (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metalaxyl and metalaxyl-M (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metamitron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methacrifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metil-S-demeton/metil-S-demton sulfon/metil-S-oksidemeton (pojedinačno ili u kombinaciji, izraženo kao metil-S-demeton)	< 0,003	< 0,006	**SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Metaflumizone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methamidophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methidathion (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methiocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methomyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxychlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Methoxyfenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metobromuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metolachlor (metolachlor including other mixtures of constituent isomers including S-metolachlor (sum of isomers))	< 0,01	< 0,01	SRPS EN 15662:2018
Metrafenone (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Metribuzin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Mevinphos (sum of cis- and trans-isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Molinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monocrotophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Monolinuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Myclobutanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Naled (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Napropamide (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Neburon (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nicosulfuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitenpyram (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nitrofen (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Norflurazon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Novaluron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Nuarimol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Omethoate (LC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Oxadixyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Oxamyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paclobutrazol (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Paraoxon-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Parathion-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pebulate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Penconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pencycuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pendimethalin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Permethrin (sum of cis- and trans- isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenamifos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenothrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phenthoate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phorate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosalone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosmet (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phosphamidon (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Phoxim (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Picoxystrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Piperonylbutoxide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimicarb desmethyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-ethyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pirimiphos-methyl (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prochloraz (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Procymidone (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Profenofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Promecarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prometon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propachlor (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propamocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propanil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propargite (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propetamphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propham (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propiconazole (sum of isomers) (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Propoxur (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Propyzamide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Proquinazid (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prosulfocarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothioconazol-desthio (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Prothiofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pymetrozine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyraclostrobin (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyracarbolid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrethrins (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyridaben (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyrimethanil (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Pyriproxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quinoxyfen (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Quintozene (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Resmethrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Rotenone (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Secbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Siduron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simazine (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Simetryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spinosad (spinosad, sum of spinosyn A and spinosyn D) (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirodiclofen (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiromesifen (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spirotetramat (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Spiroxamine (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulfotep (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Sulprofos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tau-Fluvalinate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuconazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
Tebufenozide (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebufenpyrad (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tebuthiuron (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tecnazene (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Teflubenzuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tefluthrin (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbacil (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbufos (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfone (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbufos-sulfoxide (GC-MS/MS)	< 0,003	< 0,003	SRPS EN 15662:2018
Terbumeton (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbuthylazine (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Terbutryn (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetrachlorvinphos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetraconazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetradifon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tetramethrin (sum of isomers) (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiabendazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiacloprid (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiamethoxam (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiencarbazone-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thifensulfuron-methyl	< 0,01	< 0,01	SRPS EN 15662:2018
Thiobencarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiodicarb (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Thiophanate-methyl (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolclofos-methyl (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tolylfluanid (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Transfluthrin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimefon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triadimenol (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triazophos (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trichlorfon (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Tricyclazole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifloxystrobin	< 0,01	< 0,01	SRPS EN 15662:2018

Parametri predviđeni Pravilnikom o zdravstvenoj ispravnosti dijetetskih proizvoda (Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022)			
Parametar	Vrednost	Referentna vrednost	Oznaka metode
(GC-MS/MS;LC-MS/MS)			
Triflumizole (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triflumuron (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Trifluralin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Triticonazole (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Valifenalate (LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vernolate (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Vinclozolin (GC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Zoxamide (GC-MS/MS;LC-MS/MS)	< 0,01	< 0,01	SRPS EN 15662:2018
Dithiocarbamates (dIthiocarbamates expressed as CS2, including maneb, mancozeb, metiram, propineb, thiram and ziram)	< 0,01	< 0,01	***SRPS EN 12396-2:2008
MIKOTOKSINI µg/kg			
Aflatoksin B1	< 0,05	< 0,1	VDM 0013
Ohratoksin A	< 0,25	< 0,5	VDM 0014
Zearalenon	< 5	< 20	VDM 0227
Deoksinivalenol	< 25	< 200	VDM 0246
Patulin	< 4	< 10	VDM 0015
Fumonizini (zbir B1 i B2)	< 25	< 200	VDM 0289
ORGANSKI PARAMETRI µg/kg			
Benzo(a)piren	< 0,2	< 1	VDM 0021
Dostavljeni uzorak odgovara sa aspekta ispitivanih parametara čl. 50 i čl. 51 Pravilnika Sl. glasnik RS, br. 45/2010, 27/2011, 50/2012, 21/2015, 75/2015, 7/2017, 103/2018, 45/2022.			

(approval, signatures and names redacted)

Sample code: L1 (product name, lot and sample ID redacted)

modifikacija u delu područja primene (tačka 1.), i u delu kalibracione krive (tačka 7.3.)
VDM 0295 Journal of Chromatography A, 1081 (2005) 145–155, Residue determination of glyphosate, glufosinate and aminomethylphosphonic acid in water and soil samples by liquid chromatography coupled to electrospray tandem mass spectrometry, María Ibáñez, Óscar J.Pozo, Juan V.Sancho, Francisco J.López, Félix Hernández
VDM 0297 EURL-SRM – Analytical Method Report, Analysis of Dithianon in Food of Plant Origin using acidified QuEChERS and LC-MS/MS, Version 2 (last update: 29.04.16)
VDM 0299 Modified QuEChERS-Method for the Analysis of Chlorothalonil in Fruits and Vegetables, EU Reference Laboratories for Residues of Pesticides SRM 10
**parameter van obima akreditacije
***rezultat van akreditovanog opsega merenja

Kraj Izveštaja

Sample code: L2 (product name, lot and sample ID redacted)

(2009) 716–720. LCTech, Sample extraction procedure for Ochratoxin A.
VDM 0021 EPA Method 610: Polynuclear Aromatic Hydrocarbons. Priprema prema modifikovanoj metodi:AOAC Official Method 973.30 Polycyclic Aromatic Hydrocarbons and Benzo(a)pyrene in Food, Spectrophotometric method, 1998. BS EN 15662:2008 Foods of plant origin - Determination of pesticide residues using GC/MS and/or LCMS/MS following acetonitrile extraction/partitioning and clean-up by dispersive SPE. QuEChERSmethod.
VDM 0227 SRPS EN 15850:2012 Prehrambeni proizvodi – Određivanje zearalenona u hrani za bebe na bazi kukuruza, u ječmenom brašnu, kukuruznom brašnu, palenti, pšeničnom brašnu i u hrani na bazi žita za odojčad i malu decu – HPLC metoda sa prečišćavanjem na imunoafinitetnoj koloni i fluorescentnom detekcijom– modifikacija u delu područja primene (tačka 1.), u delu reagenasa-mobilne faze (tačka 4.16.) i delu opsega kalibracione krive (tačka 7.3)
VDM 0244 SRPS EN ISO 14501:2010 Mleko i mleko u prahu – Određivanje sadržaja aflatoksina M1 – Prečišćavanje imunoafinitetnom hromatografijom i određivanje tečnom hromatografijom visoke performanse - modifikacija u delu područja primene (tačka 1.), i u delu kalibracione krive (tačka 7.3.)
VDM 0246 SRPS EN 15791:2011 Hrana za životinje — Određivanje deoksinivalenola u hrani za životinje — HPLC metoda sa UV detekcijom i prečišćavanjem na imunoafinitetnoj koloni izmenjen u delu oblasti ispitivanja i opsegu određivanja
VDM 0289 C.Paepens et al., Development of a liquid chromatography/tandem mass spectrometry method for the quantification of fumonisin B1, B2 and B3 in cornflakes, Rapid Commun. Mass Spectrom, Vol 19, (2005) 2021-2029
VDM 0295 Journal of Chromatography A, 1081 (2005) 145–155, Residue determination of glyphosate, glufosinate and aminomethylphosphonic acid in water and soil samples by liquid chromatography coupled to electrospray tandem mass spectrometry, Maria Ibáñez, Óscar J.Pozo, Juan V.Sancho, Francisco J.López, Félix Hernández
VDM 0297 EURL-SRM – Analytical Method Report, Analysis of Dithianon in Food of Plant Origin using acidified QuEChERS and LC-MS/MS, Version 2 (last update: 29.04.16)
VDM 0299 Modified QuEChERS-Method for the Analysis of Chlorothalonil in Fruits and Vegetables, EU Reference Laboratories for Residues of Pesticides SRM 10
**parameter van obima akreditacije
***rezultat van akreditovanog opsega merenja

Kraj Izveštaja

Sample code: L3 (product name, lot and sample ID redacted)

VDM [illegible] Control, 20 (2009) 716–720. LCTech, Sample extraction procedure for Ochratoxin A.
VDM 0015 MaryW. Trucksessand Yifeng Tang, Solid Phase Extraction Methodfor Patulinin Apple Juiceand Unfiltered Apple Juice, MycotoxinProtocols, Vol.157 (2001), pgs. 205-213.
VDM 0021 EPA Method 610: Polynuclear Aromatic Hydrocarbons. Priprema prema modifikovanoj metodи:AOAC Official Method 973.30 Polycyclic Aromatic Hydrocarbons and Benzo(a)pyrene in Food, Spectrophotometric method, 1998. BS EN 15662:2008 Foods of plant origin - Determination of pesticide residues using GC/MS and/or LCMS/MS following acetonitrile extraction/partitioning and clean-up by dispersive SPE. QuEChERSmethod.
VDM 0227 SRPS EN 15850:2012 Prehrambeni proizvodi – Određivanje zearalenona u hrani za bebe na bazi kukuruza, u ječmenom brašnu, kukuruznom brašnu, palenti, pšeničnom brašnu i u hrani na bazi žita za odojčad i malu decu – HPLC metoda sa prečišćavanjem na imunoafinitetnoj koloni i fluorescentnom detekcijom– modifikacija u delu područja primene (tačka 1.), u delu reagenasa-mobilne faze (tačka 4.16.) i delu opsega kalibracione krive (tačka 7.3)
VDM 0246 SRPS EN 15791:2011 Hrana za životinje — Određivanje deoksinivalenola u hrani za životinje — HPLC metoda sa UV detekcijom i prečišćavanjem na imunoafinitetnoj koloni izmenjen u delu oblasti ispitivanja i opsegu određivanja
VDM 0289 C.Paepens et al., Development of a liquid chromatography/tandem mass spectrometry method for the quantification of fumonisin B1, B2 and B3 in cornflakes, Rapid Commun. Mass Spectrom, Vol 19, (2005) 2021-2029
VDM 0295 Journal of Chromatography A, 1081 (2005) 145–155, Residue determination of glyphosate, glufosinate and aminomethylphosphonic acid in water and soil samples by liquid chromatography coupled to electrospray tandem mass spectrometry, María Ibáñez, Óscar J.Pozo, Juan V.Sancho, Francisco J.López, Félix Hernández
VDM 0297 EURL-SRM – Analytical Method Report, Analysis of Dithianon in Food of Plant Origin using acidified QuEChERS and LC-MS/MS, Version 2 (last update: 29.04.16)
VDM 0299 Modified QuEChERS-Method for the Analysis of Chlorothalonil in Fruits and Vegetables, EU Reference Laboratories for Residues of Pesticides SRM 10

** parametar van obima akreditacije
*** rezultat van akreditovanog opsega merenja

Kraj Izveštaja