

Supp Table 1. Operative time stratified by PGY level

	Resident (n = 48)	Operative time (IQR)	p-value
PGY Level (%)			< 0.01 †
2	1 (2.1)	131.0	
3	33 (68.8)	76.6 (60.0 – 82.0)	
4	3 (6.3)	125.0 (89.0 – 190.0)	
5	11 (22.9)	107.8 (81.0 – 133.0)	

Supp Table 2. Breakdown of device related complications

	Total (n=251)	RNFA (n = 203)	Resident (n = 48)
Device complication (%)	29 (11.6)	23 (11.3)	6 (12.5)
Erosion	2 (0.8)	1 (0.5)	1 (2.1)
Perforation	4 (1.6)	3 (1.5)	1 (2.1)
Malposition	13 (5.2)	10 (4.9)	3 (6.3)
Malleable placed	1 (0.4)	0	1 (2.1)
Patient dissatisfaction	4 (1.6)	4 (2.0)	0
Malfunction	5 (2.0)	5 (2.5)	0

Supp Table 3. Operative time stratified by new vs revision IPP

	RNFA	n	Resident	n	p-Value
Operative Time (min, IQR)					
New IPP	66.5 (57.0 – 74.0)	175	76.2 (61.5 - 82.0)	33	< 0.01 †
Redo IPP	101.9 (79.5 – 125.8)	28	113.60 (86.0 – 133.0)	15	0.41